

Metadata

| | |
|-------------------------------------|---|
| Standardized Instrument Name | WETLabs ECO Triplet sensor |
| Scheme URI | http://vocab.nerc.ac.uk/collection/L22/current/TOOL0674/2/ |
| Term URI | SDN:L22::TOOL0674 |
| Instrument DOI | |
| Instrument Unique ID | BBFL2WB-1442 |
| ID Type | Serial Number |
| Model No. | 1442 |
| Instrument Type | Baseline |
| Description | Ecotriplet 1442 was deployed on GL_LWH_M |
| Manufacturer | Wetlabs |
| Manufacturer type | Organizational |
| Notes | |
| Sensor Details | |
| Sensor Details 1 | |
| Instrument Sensor Name | ECO Phycocyanin (PC) Fluorometer |
| Instrument Sensor Serial No. | BBFL2WB-1442 |
| Sensor Range | 0-175 |
| Sensor Sensitivity | 0.086 |
| Sensor Units | ppb |
| Last Calibration Date | 2019-02-15 |
| Sensor Details 2 | |
| Instrument Sensor Name | ECO Chlorophyll (Chl) Flurometer |
| Instrument Sensor Serial No. | BBFL2WB-1442 |
| Sensor Range | 0-30; 0-50 |
| Sensor Sensitivity | 0.015 |
| Sensor Units | ug/l |
| Last Calibration Date | 2019-02-15 |
| Sensor Details 3 | |
| Instrument Sensor Name | Scattering Meter |

| | |
|-------------------------------------|--------------|
| Instrument Sensor Serial No. | BBFL2WB-1442 |
| Sensor Range | 0-5; 0-3 |
| Sensor Sensitivity | 0.003; 0.002 |
| Sensor Units | m-1 |
| Last Calibration Date | 2019-02-15 |